



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/748,811	12/30/2003	Kristine B. Fuimaono	51640/AW/W112	7196
23363 7590 11/15/2007 CHRISTIE, PARKER & HALE, LLP PO BOX 7068 PASADENA, CA 91109-7068			EXAMINER BOUCHELLE, LAURA A	
			ART UNIT 3763	PAPER NUMBER
			MAIL DATE 11/15/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/748,811

Applicant(s)

FUIMAONO ET AL.

Examiner

Laura A. Bouchelle

Art Unit

3763

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 04 September 2007.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-27 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-27 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☒ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- ☐ Notice of Informal Patent Application
- ☐ Other: _____

DETAILED ACTION

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(a) the invention was known or used by others in this country, or patented or described in a printed publication in this or a foreign country, before the invention thereof by the applicant for a patent.

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

1. Claims 1-6, 7-9, 13-16, 19-23, 24-2 are rejected under 35 U.S.C. 102(e) or (a) as being anticipated by Haissaguerre et al (US 6068629). Haissaguerre discloses a system for tissue ablation comprising a rigid probe body 6, an electrode body 34, 36 having at least one irrigation opening (Col. 12, lines 47-51), the electrode being attached to the probe at a point along the elongate body of the electrode between the first and second ends, and the distal end of the probe body and the electrode form an angle of approximately 90 degrees. The device comprises an infusion tube 75 for introducing fluid into the inner cavity. The device comprises a handle 4 and a stiffening wire 66. There may be an insulating sheath proximal to the electrode body (Col. 12, lines 47-48).

Claim Rejections - 35 USC § 103

2. The text of those sections of Title 35, U.S. Code not included in this action can be found in a prior Office action.

3. Claims 10-12, 17, 18 are rejected under 35 U.S.C. 103(a) as being unpatentable over Haissaguerre. Claims 10-12 differ from Haissaguerre in calling for the probe to have a certain length. Claims 17, 18 differ in calling for the electrode to be of a certain size. Haissaguerre discloses that the device is for mapping and ablating tissue within the human heart. Where the only difference between the prior art and the claims was a recitation of relative dimensions of the claimed device and a device having the claimed relative dimensions would not perform differently than the prior art device, the claimed device was not patentably distinct from the prior art device. See MPEP 2144.04. Since both applicant's invention and the device of Haissaguerre are for use within the heart of a human, it would be expected that the device of Haissaguerre would perform well with the dimensions as claimed by applicant.

Response to Arguments

4. Applicant's arguments, see page 2, filed 9/4/07, with respect to the rejection(s) of claim(s) 1-8, 13-23 under Tu have been fully considered and are persuasive. Therefore, the rejection has been withdrawn. However, upon further consideration, a new ground(s) of rejection is made in view of Haissaguerre as above.

Art Unit: 3763

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Laura A. Bouchelle whose telephone number is 571-272-2125. The examiner can normally be reached on Monday-Friday 8-4.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nicholas Lucchesi can be reached on 517-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



NICHOLAS D. LUCCHESI
SUPERVISOR/PATENT EXAMINER
TECHNICAL CENTER 3700

Laura A Bouchelle
Examiner
Art Unit 3763

